


TSI SidePak™ SP330/350 Quick Reference Card



- Press to power ON/OFF.
- Press and hold for 3 seconds to power off.
- Press to return to previous menu or screen.
- to scroll through menu selections or change settings.

• to select/enter value.

Tamper Protection:
Press and hold and to lock or unlock keypad.

Display Reading
Model SP330
S/N xxxxxx
Version X
Battery xx%
ELAPSED
TIME xxx min.
READY MODE
Enter = Start

Auto Cycle

To start pump

Flow Mode
ELAPSED
Time xxx min.
BATT xxx min.

Auto Cycle

To enter setup menu

To Scroll through menu

Set up Menu
Flow Set PNT
Clear Data
Reset All
Set Program
Run Program

Pump will start
• Let pump run 2 min.
• Connect Calibrator
• Use keys to increase or decrease flow rate to desired reading on calibrator.
• Press to accept new flow rate.
• Instrument will display new estimated run time and return to ready mode.

"ENTER = CLEAR DATA"
 to execute clear Data Function

"ENTER = RESET FLOW"
to execute reset all functions

TOTAL TIME
xxx min. (5-999 min.)
• Use to select run time.
• Press to execute new run time

Delay
xxx Min (1-999 min.)
• Use to select delay start time
• Instrument returns to ready mode

• Pump will turn on and display
PROGRAM MODE
Batt xxx min.
TOTAL SAMPLE TIME
x min.
REMAINING Time
xxx min.

OR

• Pump will display
PUMP START
IN xxx min.

• If programmed to start after a delay



TTT Environmental Instruments and Supplies

The preferred source for instrument Rentals, Sales, Service and Supplies!

4201 B Street
Anchorage, AK 99503
(907) 770-9041
www.tttenviro.com

25420 74th Avenue S.
Kent, WA 98032
(253) 373-9041
info@tttenviro.com